



**Certificate of Conformance**

**ISO 16840-3 Strength for postural support devices**

As part of its risk management, Bodypoint has verified the conformance of its products in accordance with the requirements of **ISO Wheelchairs Seating – Part 3: Determination of static, impact and repetitive load strengths for postural support devices**. Testing was performed by Bodypoint in its test lab in Seattle, WA.

**We certify that the following postural supports meet the requirements for strength:**

<u>Article Nos.</u>	<u>Product Family</u>
NB200-series	Non-Padded Hip Belts, 2-pt.
HB200-series	Padded Hip Belts, 2-pt.
HB400-series	Padded Hip Belts, 4-pt.
EB200-series	Evoflex Hip Belts
LB200-series	Velcro-Compatible Padded Hip Belts, 2-pt.
LB400-series	Velcro-Compatible Padded Hip Belts, 4-pt.
SH120-series	Monoflex Chest Supports
SH210/220-series	H-style Harnesses
SH230/235/240-series	Trimline Harnesses
SH250/260/290-series	PivotFit Harnesses, Standard
SH270/280/290-series	PivotFit Harnesses, Dynamic
SH300-series	Stayflex Chest Supports
SH102	Chest Belt

Note:

Copies of test documentation are available on request.

Signed Matthew Kosh Date June 6, 2019

Matthew Kosh  
President  
[matthew@bodypoint.com](mailto:matthew@bodypoint.com)

Bodypoint, Inc. 558 First Avenue S., #300, Seattle, Washington 98104 USA